



*Attached Papers
Note Data Sheet
Attached - JRL
7-2-79*
Flint Osteopathic Hospital

A NONPROFIT CORPORATION / 3921 Beecher Road / Flint Michigan 48502 / Phone 762-4000

A.B. Davis, Chief
Fuel, Facility, and Materials Safety Branch
U.S.N.R.C.
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Dear Sir:

This is our response to your letter of June 12 1979, which listed apparent noncompliant activities as noted by Mr. Richard L Hague during his May 21 1979, inspection of our facility. Please refer to Appendix A of your letter for the following references.

Item 1: On May 31 1979, we held a meeting of the Medical Isotopes committee. One topic of discussion was the required meeting frequency. Notices of quarterly meetings will be sent to all member of the committee.

Item 2: Daily response and constancy checks of the performance of our dose calibrator were started June 1 1979. Further, we have ordered another Cobalt-57 source of the same geometry and construction as the one currently in use. This will afford two reference point of approximately 0.6 and 5.0 millicuries. Result of these checks are being logged.

Item 3: Our "Cutie Pie" type survey meter, Texas Nuclear Models 2593, 2594 and 2595 were sent to the manufacturer for preventative maintenance and calibration on May 22 1979. The units were received yesterday, June 27 1979. (See attached.)

Item 4: A procedure for facility shutdown at shift end is being developed. Within this procedure is the requirement for site and waste surveys, and recording of their results. We expect to implement this within 30 days or less.

Item 5: Notification of the location of 10 CFI parts 19 and 20 was added to the revised copy of NRC-3 supplied by Mr. Hague, and was duly posted on May 22 1979.

Please feel free to contact me should there be any questions.

Sincerely,

Joseph P. Ohlmacher

Joseph P. Ohlmacher
Health Physicist
ENC.

7909130 294

~~7909130~~

cc: Dr. Monsour
Dr. Stratton
file

JO:ps

329094

JUL 2 1979

W. DALE FERGUSON
President

BOARD OF TRUSTEES

Chairman
HOWARD A. HICKOK
Gransden-Hall

Vice Chairman
GALE R. POTTER
Chevrolet Motor Division, Retired

Secretary
JUDGE HAROLD H. BOBIER

Treasurer
W. DALE BACHTEL
AC Spark Plug, Retired

DR. MYRTLE BLACK
Mott Foundation, Retired

GERALD E. PIESKO, D.O.
Medical Staff

JOHN S. STOUT
Buick Motor Division

GERALD J. WINTER
Chevrolet Flint Assembly

PAUL R. MURPHY
General Motors Parts Division

EDWIN E. LUKE
A.G. Edwards & Sons, INC.

JAMES A. HAYES
Complete Design Service

MRS. MARGARET G. STEWART
Retired, Metropolitan Director
Girl Scouts of Greater N.Y.

LEO SEIDE
Auto City Distributors, INC.

GERALD T. MEIER
AC Spark Plug

→
L
L

Texas Nuclear
Division

Ramsey Engineering Company
Box 9267
Austin, Texas 78766 USA
Telephone (512) 836-0801
Telex 77-6413

CERTIFICATE OF INSTRUMENT CALIBRATION

This is to certify that the product described below was calibrated on the date shown below by a qualified Service Technician of the Texas Nuclear Division.

MODEL NO: 2572 SERIAL NO: 711

CALIBRATION DATA:

Calibrated & checked on all ranges against a calibrated 60 mci kr-85 source. All ranges agree +10%.

DATE: 6-25-79 NEXT DUE: 6-25-80

SIGNATURE: Mike Forrest

This is our certification to you that your product was calibrated within three (3) days prior to its shipment.

TEXAS NUCLEAR DIVISION

SERVICE DIVISION

923095