

Appendix A

NOTICE OF VIOLATION

Wisconsin Industrial Testing
Company

License No. 48-17480-01

Based on the inspection conducted on July 9, 1979, it appears that one of your activities was in noncompliance with NRC requirements, as noted below. This item is a deficiency.

Condition 18 of your license requires that licensed material be possessed and used in accordance with statements, representations, and procedures contained in application dated July 25, 1979. Section BC-1 of this application states that survey instruments in use will be calibrated by one of two specified firms.

Contrary to this requirement, as of the date of this inspection, your survey instruments were being calibrated by a firm other than one of the two specified in your license application.

751037

790920 146