

LICENSEE EVENT REPORT

CONTROL BLOCK: (PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

01 N C B E P 1 2 0 0 0 0 0 0 0 0 3 4 1 1 1 1 1 4 5

CON'T 01 REPORT SOURCE L 6 0 5 0 0 3 2 5 7 1 1 1 9 7 9 8 1 2 1 1 7 9 9

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES 10
02 During normal plant operation, the operator noticed that the containment atmospheric
03 monitor, CAC-AQH-1263, had failed downscale on the oxygen channel.
04
05
06
07

08 Technical Specification 3.6.6.4, 6.9.1.9b

09 SYSTEM CODE SE 11 CAUSE CODE E 12 CAUSE SUBCODE G 13 COMPONENT CODE INSTRU 14 COMP. SUBCODE X 15 VALVE SUBCODE Z 16
17 LER/RO REPORT NUMBER 7 9 EVENT YEAR 7 9 SHUTDOWN METHOD Z 21 HOURS 0 0 0 22 ATTACHMENT SUBMITTED N 23 NPRD-4 FORM SUB. Y 24 PRIME COMP. SUPPLIER A 25 COMPONENT MANUFACTURER B 1 3 5 26

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS 27
10 The oxygen analyzer had drifted out of calibration for no known reason. The
11 instrument was recalibrated and returned to service satisfactorily. Engineering is
12 now investigating other monitoring systems to determine the possibilities for
13 replacement.
14

15 FACILITY STATUS E 28 % POWER 0 8 4 29 OTHER STATUS N/A 30 METHOD OF DISCOVERY P 31 DISCOVERY DESCRIPTION Operator surveillance 32

16 ACTIVITY CONTENT Z 33 Z 34 AMOUNT OF ACTIVITY N/A 35 LOCATION OF RELEASE N/A 36

17 PERSONNEL EXPOSURES 0 0 0 37 Z 38 DESCRIPTION N/A 39

18 PERSONNEL INJURIES 0 0 0 40 DESCRIPTION N/A 41

19 LOSS OF OR DAMAGE TO FACILITY TYPE Z 42 DESCRIPTION N/A 43

20 PUBLICITY ISSUED N 44 DESCRIPTION N/A 45

NAME OF PREPARER A. C. Tollison, Jr. PHONE: 919-457-9521

7 912180 532

POOR ORIGINAL

i609 226