APPENDIX A

NOTICE OF VIOLATION

University of Delaware

Docket Nos. 70-[illegible]09
50-98

Based on the results of an NRC inspection conducted on November 1, 1978, it appears that certain of your activities were not conducted in full compliance with the conditions of your licenses as indicated below. Item A is an infraction and Item B is a deficiency.

A. Condition 8 of License No. SNM-656 authorizes the use of SNM in accordance with statements, representations and procedures contained in letters dated September 21, 1973, and June, 7, 1974. The September 21, 1973, letter states, in part, that three specified radiation protection instruments will be available, calibration of these instruments will be accomplished quarterly in accordance with instrument manufacturers instructions, and a permanent record of results will be kept.

Contrary to these requirements, no calibrations were performed during the eleven months preceding November 1, 1978.

B. The Technical Specifications incorporated into License No. R-43 required the submission of an annual operating report to the Director, NRC Regional Office I, by March 1 of each year to include eight specified topics.

Contrary to this requirement the annual operating reports due by March 1 of 1977 and March 1 of 1978 were not submitted.

7902089100