

Appendix A

NOTICE OF VIOLATION

University of Michigan

Docket No. 50-2

Based on the inspection conducted on September 6-8, 1978, it appears that certain of your activities were in noncompliance with NRC requirements, as noted below. The noncompliance item is a deficiency.

Technical Specification 6.5.A.3 requires that tests, checks, and measurements documenting compliance with surveillance requirements be prepared and retained by the licensee for at least five years.

Contrary to the above and Technical Specification 4.7.A, there was no documentation to confirm that the gamma spectrometry equipment used to measure liquid waste tank contents had been calibrated since March 1977.

4/
78110900 ~~78~~