



VICTOREEN, INC. • 10101 WOODLAND AVE., CLEVELAND, OHIO 44104  
PHONE: (216) 795-8200 • TWX (810) 421-8287

September 22, 1980

U. S. Nuclear Regulatory Commission  
Region III, Office of Inspection and Enforcement  
799 Roosevelt Road  
Glen Ellyn, Illinois 60137

PRINCIPAL STAFF			
DIR		SAFE	
D/D		ADM	
A/D		PAO	
✓ FFGMS		SLO	
RC&ES			
RC&NS		FILE	✗

Gentlemen:

Pursuant to condition 13 C of Byproduct Material License no. 34-00486-04, we are filing the following leaking source report:

Source vendor: ICN Corp.  
2727 Campus Drive  
Irvine, California 92715

Source type: ICN model 312, 0.5mCi of Sr-90

Number of sources involved: 54 sources dated from 5/79 to 6/80.

Number found leaking: 11 sources: removable contamination ranged from 0.004 to more than 0.05 microcuries.

Equipment: Nuclear Associates Minimonitor 125, model 05-572

Sensitivity: 5000 CPM per 0.005 microcuries

Corrective action: All 54 sources\* are being returned to Mr. Layport of ICN for examination and replacement.

\*Since these sources were all leak tested upon arrival at Victoreen and found to be not leaking, it is felt that a generic problem may exist.

Sincerely,

John C. Miller  
Radiation Safety Officer

JCM/djg

8012160

236



SEP 24 1980