

Detroit Testing Laboratory, Inc.

FULL SERVICE EVALUATION CENTER



8720 NORTHEAST AVENUE OAK PARK, MICHIGAN 48237 (313) 398-2100 TWX 810-2324847

November 19, 1980

United States
Nuclear Regulatory Commission
Region 111
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Attn: Mr. James G. Keppler

Subject: I E Bulletin 80-22: Automation Industries, Model 200-520-008
Sealed Sources Connectors.

Dear Sir:

In regards to the above bulletin, Detroit Testing Laboratory does not have and has never utilized the above mentioned Automation Industries, Model 200-520-008 Sealed Source Connectors.

Detroit Testing Laboratory utilizes IR 192 Gamma Century, Model S.A. Cameras under our license number 21-18302-01. The Gamma Century Cameras are equipped with the new Saf-T-Key coupling, which is virtually impossible to disconnect during use due to its design.

Based on our experience with and upon our operation and emergency procedures with the above equipment, we feel that no further action is required by our company as long as we continued to utilize the Gamma Century Cameras.

Thank you for your continual response and bulletins which aid companies like our selves, in being advised of possible safety problems as well as keeping up with the latest requirements.

Sincerely,
DETROIT TESTING LABORATORY, INC.

James L. Nelson
Jerry L. Nelson, Manager
Non-Destructive Testing
Radiation Safety Officer

JLN/dm

C.C. Director
Division of Fuel Facility and Materials Safety Inspection
Office of Inspection and Enforcement
U.S. Nuclear Regulatory Commission
Washington, D.C. 20555

US NRC
INSPECTION SERVICES
BRANCH

NOV 26 2 59 PM '80

8012080365

PT 9

NOV 24 1980