

JLS **SHEPHERD** *and Associates*

740 Salem Street, Glendale, California 91203

• 213/245-0187

Irradiation & Calibration Equipment

• Lead Shielding

• Nuclear Applications

October 2, 1985

Nuclear Regulatory Commission
Office of Inspection & Enforcement
Mail Stop EWW-359
Washington, D. C. 20555

Attention: Mr. James Wigginton

RE: Series 28 Calibrator
with PT-1 Time Option
Service Advisory

Dear Mr. Wigginton:

Although we have several hundred manually operated Series 28 Calibrators in service, only eight of these units (see attached list of users) are equipped with the PT-1 Time Option which incorporates the spring loaded source ON latch as covered in the Service Advisory. Other units with the PT-1 Time Option, made prior to this design change, did not incorporate the swinging latch design and were not covered by the Service Advisory.

Sincerely,


J. L. Shepherd

JLS:jlf
ENCL: User List

8510250301 851021
IE ADDCK 05000254
CF

USER LIST
SERIES 28 CALIBRATOR
WITH
PRESET TIMER OPTION

1. Vern Chaney
Commonwealth Edison Company
Mazon EOF
960 N. Route #47
Morris, Illinois 60450
2. M. P. Horne
Toledo Edison Company
Davis-Besse Station Storeroom
5501 North SR 2,
Oak Harbor, Ohio 43449
3. Paul Weeden
Arizona Radiation Regulatory Agency
925 S. 52nd Street, Suite 2,
Tempe, Arizona 85281
4. M. Vendel
Commonwealth Edison Company
Dresden Nuclear Station
RR #1,
Morris, Illinois 60450
5. Jack Kiester
Commonwealth Edison Company
Byron Nuclear Station
German Church Road,
Byron, Illinois 61010
6. G. Adamaitis
Commonwealth Edison Company
Zion Nuclear Station
101 Shilon Boulevard,
Zion, Illinois 60099
7. G. Adamaitis
Commonwealth Edison Company
Quad Cities Nuclear Station
22710 206th Avenue North,
Cordova, Illinois 61242
8. R. Grehl
Consolidated Edison Company
Indian Point Station
Broadway & Bleakley
Buchanan, New York 10511