NRC FORM 366 (7-77)

U. S. NUCLEAR REGULATORY COMMISSION

LICENSEE EVENT REPORT

CONTROL BLOCK: | | | | | | | (1) (PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

01	LICENSEE CODE: N C B E P 1 (2); LICENSE NUMBER: 0 0 - 0 0 0 0 0 - 0 0 (3); LICENSE TYPE: 4 1 1 1 1 (4); CAT: (5)
CON'T 01	REPORT SOURCE: L (6); DOCKET NUMBER: 0 5 0 - 0 3 2 5 (7); EVENT DATE: 0 3 0 8 8 0 (8); REPORT DATE: 0 3 2 5 8 0 (9)

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES (10)

During normal plant operation, the control operator noticed that the oxygen channel of containment atmospheric monitor CAC-AT-1263-2 experienced a 1% step change. The CAC-AT-1259-2 monitor showed no change in concentration. This monitor has experienced several step changes during the past several months.

Technical Specifications 3.3.5.3, 3.6.6.4, 6.9.1.9b

SYSTEM CODE	CAUSE CODE	CAUSE SUBCODE	COMPONENT CODE	COMP. SUBCODE	VALVE SUBCODE
S D (11)	E (12)	A (13)	I N S T R U (14)	X (15)	Z (16)

(17) LER/RO REPORT NUMBER	EVENT YEAR		SEQUENTIAL REPORT NO.		OCCURRENCE CODE	REPORT TYPE		REVISION NO.
	8 0	—	0 0 4	/	0 3	L	—	0

ACTION TAKEN	FUTURE ACTION	EFFECT ON PLANT	SHUTDOWN METHOD	HOURS (22)	ATTACHMENT SUBMITTED	NPRD-4 FORM SUB.	PRIME COMP. SUPPLIER	COMPONENT MANUFACTURER
X (18)	X (19)	Z (20)	Z (21)	0 0 0 0	N (23)	Y (24)	A (25)	B 1 3 5 (26)

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS (27)

Past module replacement on this instrument has failed to correct the step increase problem; therefore, it is believed that a wiring or connection problem exists. A Technical Representative from Beckman is on site to assist in repairing this monitor. A calibration always returns these instruments to normal and they are calibrated every two weeks.

FACILITY STATUS	% POWER	OTHER STATUS (30)	METHOD OF DISCOVERY	DISCOVERY DESCRIPTION (32)
E (28)	0 9 8 (29)	NA	A (31)	Operator Surveillance

ACTIVITY RELEASED	CONTENT OF RELEASE	AMOUNT OF ACTIVITY (35)	LOCATION OF RELEASE (36)
Z (33)	Z (34)	NA	NA

PERSONNEL EXPOSURES

NUMBER	TYPE	DESCRIPTION (39)
0 0 0 (37)	Z (38)	NA

PERSONNEL INJURIES

NUMBER	DESCRIPTION (41)
0 0 0 (40)	NA

LOSS OF OR DAMAGE TO FACILITY (43)

TYPE	DESCRIPTION
Z (42)	NA

PUBLICITY

ISSUED	DESCRIPTION (45)	NRC USE ONLY
N (44)	NA	

NAME OF PREPARER A. C. Tollison, Jr. PHONE: 919-457-9521

8004010 390