

8005070 139



IN VIVO IMAGING
AT YOUR FACILITIES

3304 Commercial Avenue
Northbrook, Illinois 60062
312/564-3330
Chicago 312/273-2525

ISOTOPE MEASUREMENTS LABORATORIES, INC.

April 7, 1980

Mr. A.B. Davis, Chief
Fuel Facility & Materials Safety Branch
U.S. Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Subject: License # 12-13568-01, your letter of March 20, 1980

Dear Mr. Davis:

In accordance with your instructions submitted in subject letter we are furnishing the following information:

1. As noted in the body of your letter we have submitted our final written evaluation of a high film badge reading on one of our employees. As soon as we were notified by R.S. Landauer, Co. of the reading we had made oral and telephone investigations of the cause and determined that the reading did not represent a personnel exposure. Because of communication difficulties we did not receive written confirmation of the cause from our employee until recently. His statement and our investigation are now in our files and were forwarded to you on February 21, 1980. We have requested a change in the subject badge report to reflect an average reading for the period covered.
2. Since the visit of Ms. Nicholson and Mr. Oberg, records of receipt and transfer of the 200 mCi vials of ¹³⁷Cs have been found and are filed for ready reference in the source cabinet and are available for your inspection.
3. Since the visit of Ms. Nicholson and Mr. Oberg, leak test records for ¹³¹Cs sources stored and transferred have been located. The records were in a log of client leak test results and therefore not readily retrievable during your visit. These records have since been separated and appropriately filed in the source cabinet.

We are most embarrassed by the sited deficiencies, however, please be advised we are very cognizant of our responsibilities to the commission, our employees, and the general public. The deficiencies noted were in the unavailability of records during your visit. This deficiency has since been corrected. We welcome your visit and comments.

Very truly yours,
ISOTOPE MEASUREMENTS LABORATORIES INC.

Theodore Fields, M.S. FACR
President

TF:se

APR 11 1980