

8005080110 KANSAS GAS AND ELECTRIC COMPANY

140

GLENN L. KOESTER

April 1, 1980

Mr. U.C. Seidle, Chief Reactor Construction and Engineering Support Branch U.S. Nuclear Regulatory Commission Region IV 611 Ryan Plaza Drive Suite 1000 Arlington, Texas 76011

> Re: (a) Docket No. STN 50-482 (b) GKoester letter to WSeidle dated February 8, 1980

Subj: Tinius-Olson Concrete Cylinder Compressive Strength Testing Machine

Dear Mr. Seidle:

As indicated in my letter of February 8, 1980, (Reference (b)), we have concluded our investigation into the use of an oversize head on our Tinius-Olson compressive strength testing machine. We have concluded that the use of the oversize head has no safety significance and that Regulation 10CFR50.55(e) is not applicable. Data to support this conclusion is available at the Wolf Creek Site and may be reviewed during any future site inspection.

Yours very truly,

Alenn Louster

GLK:bb