

April 11, 1980

Veterans
Administration

8005160040



United States Nuclear Regulatory Commission
Region I
631 Park Avenue
King of Prussia, Pennsylvania 19406

Gentlemen:

Concerning the findings of inspection 80-01 of NRC licenses 08-00942-02, 08-00942-05, and SNM-1065 conducted on March 11 and March 19, 1980, filed at docket 30-1314, the following is submitted.

Administrative controls have been established to insure that the dose calibrator will be tested for linearity quarterly and after each repair.

Daily GM surveys of the nuclear medicine hot lab and clinic have been instituted.

One lead glass syringe shield is on hand and a second has been ordered for use in the preparation of ^{99m}Tc -labelled products. Their use by nuclear medicine technologists has been mandated by the station Radiation Safety Officer and the Chief of the Nuclear Medicine Service.

Sincerely,

A handwritten signature in cursive script, appearing to read 'A.A. Gavazzi'.

A.A. Gavazzi
Director

cc: James J. Smith, M.D.
Director of Nuclear Medicine Services