

Mohamed M. Sharbaky  
Region 1  
US Nuclear Commission  
475 Allendale Road  
King of Prussia, PA 19406

RE: Docket # 030-30131

Lic # 37-28074-01

Inspection # 030-30131/69-031

With respect to the above inspection, please be advised that the measurements of the reference standards used in the daily constancy test are now compared to an expected value determined by the decay of the standard used. This measurement is compared to the decay corrected activity to determine the percent error in the measurement and to ensure that this error is less than 10%.

A quarterly audit by an outside consultant is performed to ensure compliance.

Sincerely,

  
Surendra V. Pawar

8911150024 891101  
REG1 LIC30  
37-28074-01 PDC