

NOTICE OF VIOLATION

Park Plaza Radiology, Inc.

License No. 21-20320-01

As a result of the inspection conducted on October 10, 1989, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1989) (Enforcement Policy) the following violation was identified:

10 CFR 35.70(h) requires that survey records for removable contamination in each area be expressed in disintegrations per minute per 100 square centimeters.

Contrary to the above, from February 2, 1987 through October 10, 1989, the licensee had not been recording weekly removable contamination surveys in disintegrations per minute. Specifically, the licensee had been recording the results of those surveys in counts per minute rather than disintegrations per minute.

This is a Severity Level V violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Oct 23, 1989  
Dated

Roy J. Caniano  
Roy J. Caniano, Chief  
Nuclear Materials Safety  
Section 2