

OCT 11 1989

Tedia Company, Inc.  
ATTN: Mr. Craig Lange  
Post Office Box 342  
Fairfield, OH 45014

License No. 34-16769-01

Gentlemen:

This refers to a telephone contact conducted September 21, 1989. The contact was an examination of activities conducted under your license as they relate to radiation safety and to compliance with the Commission's rules and regulations and the conditions of your license. The contact consisted of a discussion with Mr. Craig Lange.

As a result of this examination of activities, regulatory concerns were noted and are specified below. They may be evaluated at an onsite inspection at your facility in the near future.

1. Regulatory Concerns: leak tests of the gas chromatograph sealed source (foil) is not being performed at the required frequency.
2. The RPO authorized on your license, Mr. Bruce BeCraft, resigned his position and Craig Longe assumed the responsibilities of RPO. These changes were made without obtaining prior NRC approval and amending your license to reflect the changes.

Bruce BeCraft, RPO of the license, resigned his position without NRC notification and Craig Longe assumed the responsibilities of RPO without an amendment being made to the materials license.

You should examine your license and NRC regulations to determine how you can correct these apparent regulatory concerns. Please submit to this office within 30 days of this letter, a description of actions taken or those that will be taken to correct the above mentioned regulatory concerns.

If you have any questions regarding this contact, you may contact us at 312-790-5500.

Sincerely,

Don Sreniawski, Chief  
Nuclear Materials Safety  
Section 1

cc: DCD/DCB (RIDS)

R111  
King/lms

*Yes*  
R111 *WJB*  
Sreniawski *10/3*

*1/0*  
**IE07**