



30-20812

ST. MARY'S MEDICAL CENTER

3700 Washington Avenue · Evansville, Indiana 47750 · (812) 479-4000

October 9, 1989

D. J. Sreniawski, Chief
Nuclear Materials Safety
US NRC Region III
799 Roosevelt Road
Glen Ellyn, IL 60137

REF: **License No. 13-03226-04**

Dear Mr. Sreniawski:

With reference to your letter of September 26, 1988 regarding the safety inspection conducted by Ms. Evelyn Matson, the following is submitted for your consideration:

Item 1

Violation was corrected in the presence of Ms. Evelyn Matson. Strontium-90 eye applicator was brought back by Alvin Korba, M.D. from Southwestern Indiana Cancer Center during the inspection. In future, no material under this license will be transferred without proper authorization.

Item 2

Since the time of inspection on August 31, 1989, a procedure has been initiated that prior to a shipment of spent Mo 99/T_c99m generators, measurements will be made to assess the contamination levels on the external surfaces of each package.

I will be glad to provide any additional information you may need in this regard.

Sincerely,

S. M. Shah

Saiyid M. Shah, Ph.D.
Radiation Safety Officer

cc: Mrs. Carol Kent
Sam Vowells
Todd Kranpitz

OCT 12 1989

IE07