



DLD/DCB

353 North Oak Street
Columbia City, Indiana 46725
Telephone 219/244-6191

WHITLEY COUNTY MEMORIAL HOSPITAL

September 19, 1989

Nuclear Regulatory Commission
Region III, 799 Roosevelt Road
Glen Ellyn, IL 60137

RE: WCMH Docket # 030-12236
License # 13-17129-01

Dear Sirs:

September 13,

This is in reply to your letter of ~~August 24~~, 1989, requesting a written reply to corrective action being taken on two violations that were found during the Nuclear Regulatory Commission inspection on July 27, 1989.

Corrective action taken on violation of 10 CFR 35.13:

On July 28, 1989, we sent an amendment to have Dr. Rausch approved as our Radiation Safety Officer. He is currently acting as our Radiation Safety Officer at our facility. The amendment was approved as of August 28, 1989 (amendment # 09). We are now in compliance with section 10 CFR 35.13.

Corrective action taken on license condition number 16 to follow appendix I of the regulatory guide 10.8 (1980):

A well counter capable of detecting 200 dPM has been obtained and has been in use as of September 15, 1989.

We should now be in compliance with both violations as of September 15, 1989. If you have any other questions or concerns, please contact me.

Sincerely,

James S. Krieg, R.T.
Radiology

JSK/glf

8910270146 890919
REG3 LIC30
13-17129-01 PNU

10 SEP 25 1989

LE07