



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION I  
631 PARK AVENUE  
KING OF PRUSSIA, PENNSYLVANIA 19406

JUN 24 1980

Docket Nos. 30-01300  
70-02749

The Griffin Hospital  
ATTN: Alfred E. Fasulo, Jr.  
Administrator  
130 Division Street  
Derby, Connecticut 06418

Gentlemen:

Subject: Combined Inspection 30-01300/80-01; 70-02749/80-01

This refers to your letter dated May 19, 1980, in response to our letter dated April 24, 1980.


Thank you for informing us of the corrective and preventive actions you documented in response to our correspondence. However, your letter has been reviewed and additional information is found to be needed.

In response to item A, you failed to indicate how you would ensure that all personnel working in your restricted area (nuclear medicine) would be properly instructed in all applicable revisions of NRC regulations and your license. You have indicated that identified surveys and dose calibrator checks are now being performed, and this is responsive to items B.2 and B.3. However, you need to assure that personnel are aware of all appropriate requirements and that all requirements are followed.

In response to item B, you have indicated that weekly wipe tests of your sealed sources are being performed and the results recorded. This test is only required to be performed at six month intervals, but must have a sensitivity of 0.005 microcuries. You have failed to provide information that the test you are now performing has the required sensitivity.

Please respond to this office within seven (7) days of receipt of this letter a written statement containing the requested information.

Sincerely,

  
John D. Kinneman, Chief  
Materials Radiological Protection  
Section

cc:  
Thomas Anastasio, Assistant Administrator

8008060184

