

JUN 16 1988

Robinson Memorial Hospital
ATTN: Donald A. Hammel, M.D.
Radiation Safety Officer
6847 North Chestnut Street
Ravenna, OH 44266

Gentlemen:

Enclosed is Amendment No. 32 renewing your NRC License No. 34-09621-01 in accordance with your request.

Please note that your quarterly surveys of sealed source and brachytherapy storage areas should be performed with a radiation measurement survey meter. Your submittal, in Attachment 10.12, states that these surveys will be performed with a radiation detection instrument. 10 CFR Part 35, Section 35.420 states in part that a licensee authorized to use byproduct material for implant therapy shall have in its possession a portable radiation measurement survey instrument capable for measuring dose rates over the range 1 millirem per hour to 1,000 millirem per hour. Also, note that we have added condition 14. to your license at the request of your consultant, to cover the equipment you need for brachytherapy sources.

Please review the enclosed document carefully and be sure that you understand all conditions. You must conduct your program involving radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, note that you must:

1. Operate in accordance with NRC regulations 10 CFR Part 19, "Notices, Instructions and Reports to Workers; Inspections," 10 CFR Part 20, "Standards for Protection Against Radiation," and other applicable regulations.
2. Possess radioactive material only in the quantity and form indicated in your license.
3. Use radioactive material only for the purpose(s) indicated in your license.
4. Notify NRC in writing of any change in mailing address.
5. Request and obtain appropriate amendment if you plan to change ownership of your organization, change locations of radioactive material, or make any other changes in your facility or program which are contrary to your license conditions or representations made in your license application and any supplemental correspondence with NRC. Any amendment request should be accompanied by the appropriate fee specified in 10 CFR Part 170.

8904120063 880616
REG3 LIC30
34-09621-01 PNU

6. Submit a complete renewal application with proper fee or termination request at least 30 days before the expiration date on your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of radioactive material after your license expires is a violation of NRC regulations.
7. Request termination of your license if you plan to permanently discontinue activities involving radioactive material prior to your expiration date.

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions and representations in your license application will result in enforcement action against you in accordance with the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

If you have any questions or require clarification of any of the above stated information, contact us at (312) 790-5625.

Sincerely,

Original Signed By
John R. Madera
Materials Licensing Section

Enclosure: Amendment No. 32

R111

Madera/bt
6/ /88