

DCD/DCB

# SOUTHWESTERN INDIANA RADIATION ONCOLOGY CENTER

Alvin Korba, M.D.  
Aly Razek, M.D.  
Thomas P. Hayes, M.D.  
Amr M. Aref, M.D.  
Shannon Lamb, M.D.

Saiyid M. Shah, Ph.D. (Physicist)  
Arnold Sorensen, B.S. (Physicist)  
John O'Donoghue (Dosimetrist)  
Kevin Maroney (Dosimetrist)

*Evansville Center  
etc*

October 3, 1989

D. J. Sreniawski, Chief  
Nuclear Materials Safety  
Section 2  
US NRC, Region III  
799 Roosevelt Road  
Glen Ellyn, IL 60137

Ref: **License Number 13-25945-01**

Dear Mr. Sreniawski:

The following is submitted with reference to your letter of September 26, 1989.

The technologists have been reminded that the NRC license condition number 14 requires that each treatment day the Gamma Med Afterloader is used, a door interlock check will be performed (copy of the memorandum is enclosed). Periodically, a member of Physics staff will check the records to ensure compliance.

Sincerely,

S. M. Shah

Saiyid M. Shah, Ph.D.  
Radiation Safety Officer

Enclosure

8910270183 891003  
REG3 LIC30  
13-25945-01 PNU

11  
2 E 07  
OCT 5 1989

# SOUTHWESTERN INDIANA RADIATION ONCOLOGY CENTER

Alvin Korba, M.D.  
Aly Razeq, M.D.  
Thomas P. Hayes, M.D.  
Amr M. Aref, M.D.  
Shannon Lamb, M.D.

Saiyid M. Shah, Ph.D. (Physicist)  
Arnold Sorensen, B.S. (Physicist)  
John O'Donoghue (Dosimetrist)  
Kevin Maroney (Dosimetrist)

TO: All Therapy Technologists  
FROM: Saiyid M. Shah, Ph.D. SMS  
DATE: October 3, 1989  
RE: Gamma Med Afterloader Door Interlock Check

As per condition 14 of our NRC license, it is required that each treatment day the Gamma Med Afterloader is used, a door interlock check will be performed.

Please make sure that the door interlock is checked and the record of the same is kept in order for us to be in compliance of NRC regulations.

Thanks for your cooperation.