

OCT 26 1989

DCS

Docket No. 030-30810  
License No. 08-28277-01  
EA 89-27

Christian E. Chinwuba, M.D.  
Southwest Imaging Center  
667 E. Street, S.W.  
Washington, D.C. 20024

Dear Dr. Chinwuba:

This will acknowledge receipt of your check for \$250.00 in payment for the civil penalty imposed by NRC in its Order dated July 7, 1989.

Concerning your response to Violation F as summarized in the Appendix to the Order, 10 CFR 35.220 requires that you have on hand a portable radiation survey instrument capable of detecting dose rates over the range of 0.1 millirem per hour to 100 millirem per hour and a portable radiation survey instrument capable of measuring dose rates over the range of 1 millirem per hour to 1000 millirem per hour. If your survey instrument is sent away for calibration or repair, it is acceptable to have on hand a "loaner" (borrowed) instrument provided that it is capable of measuring dose rates over the range specified in 10 CFR 35.220.

Your corrective actions will be examined during future inspections.

Sincerely,

**Original Signed By**  
**James Lieberman**

James Lieberman, Director  
Office of Enforcement

cc: W. Russell, RI

Distribution  
JLieberman  
JDeMedico  
DHoody, RI  
EA File  
ES File

*per K. Christopher*

OE *D*  
JDeMedico  
10/24/89

RA:RI  
WRussell  
10/19/89

OE: *D*  
JLieberman  
10/15/89

8911020104 891026  
NMSS LIC30  
08-28277-01 PNU

1/0  
IE 14