

NOV 01 1989

Docket No. 030-11944

License No. 18-16956-01

Cary Medical Center
ATTN: Lee Myles
Associate Director of Operations
MRA Box 37, Van Buren Road
Caribou, Maine 04736

Gentlemen:

Subject: Inspection No. 030-11944/89-01

This refers to your letter dated September 8, 1989, in response to our letter dated August 4, 1989.

Thank you for informing us of the corrective and preventive actions documented in your letter. These actions will be examined during a future inspection of your licensed program.

Item A of your letter states that the RSO does ensure all radiation safety activities are performed in accordance with approved procedures and a copy of your package receipt policy was attached for our review. At the time of inspection, our inspector noted that the technologist only wiped the package if a leak was suspected. Your current policy states all packages will be wiped. Please identify your method for ensuring your approved package receipt policy will be followed in the future. Please identify the date at which full compliance will be achieved. Please submit the required information to this office within thirty days of the date of this letter.

Your cooperation with us is appreciated.

Sincerely,

Original Signed By:
Mohamed M. Shanbaky

Mohamed M. Shanbaky, Chief
Nuclear Materials Safety Section A
Division of Radiation Safety
and Safeguards

cc:
Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
State of Maine

OFFICIAL RECORD COPY

RL CARY - 0001.0.0
10/30/89

8911150014 891101
REG1 LIC30
18-16956-01 PDC

RETURN ORIGINAL TO
REGION I

IE:07

bcc:
Region I Docket Room (w/concurrences)

RI:DRSS *8/28*
Joustra/Bj

10/30/89

RI:DRSS
Shanbaky
ms

10/30/89

OFFICIAL RECORD COPY

RL CARY - 0002.0.0
10/30/89