

NOTICE OF VIOLATION

Mt. Carmel Mercy Hospital  
Detroit, Michigan

License No. 21-00998-02

As a result of the inspection conducted on October 25 and 26, 1989, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1989) (Enforcement Policy) the following violation was identified:

10 CFR 32.632(f) requires full calibration measurements of teletherapy equipment be performed by the teletherapy physicist.

Contrary to the above, the full calibration performed on the Theratron 80 teletherapy unit on December 18, 1988 was performed by an individual who is not listed as the teletherapy physicist.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

November 9, 1989  
Dated

Roy J. Caniano  
Roy J. Caniano, Chief  
Nuclear Material Safety  
Section 2