

June 6, 1980

License SNM-1067  
Docket 70-1100

U. S. Nuclear Regulatory Commission  
Region I  
631 Park Avenue  
King of Prussia, PA 19406

Attention: Mr. Walter G. Martin, Chief  
Safeguards Branch

Reference: Letter from Mr. Walter G. Martin to Mr. H. V. Lichtenberger,  
dated May 22, 1980; Inspection 70-1100/80-03

Gentlemen:

This is in response to the above referenced letter in which you reported one item of noncompliance which was determined during your inspector's visit to our facility on April 28-May 1, 1980.


Paraphrasing from Appendix A, the item of noncompliance consisted of using values for inventory purposes that were outside of the calibration range for our nondestructive assay machine.

To correct the situation, the following action will be taken:

New standards will be developed for our assay machine so that the range of values measured will be within calibration limits. When the new standards are available, no values outside of the calibration range of the assay machine will be used for accountability purposes.

The new standards will be available for use by July 1, 1980.

Very truly yours,

  
H. V. Lichtenberger  
Vice President-Nuclear Fuel  
Nuclear Power Systems-Manufacturing

HVL/HTC/ssb

8007280737