

INSTRUMENT ERROR
ANALYSIS FOR FORT
CALHOUN CLRT
January 1980

Prepared for
Omaha Public Power District

By

Nuclear Services Corporation
2622 E. 21st. St.
Tulsa, OK 74114

(918) 749-2582

Prepared by: R.L. Novgrod
Reviewed by: C.B. Noland
Approved by: C.W. Rowley

July 11, 1980

OMA-0110

8008130 525

INSTRUMENT ERROR ANALYSIS

I. Instrumentation

A. Pressure Digital Readout

Inst #	PI-1, PI-2
Mfg	Mensor Corp.
Type	Bourdon Tube Capsule & Precision Pressure Gage
Range	0-100 psia
Stability	+ .002 psi
Repeatability	+ .0005 psi
Resolution	+ .001 psi
Accuracy	+ .015 psi

B. Temperature Elements

Inst #	T-001 through T-030
Mfg	Hy-Cal Engrg
Type	100 Ω Platinum RTD
Range	Calibrated for 60°, 85°, and 110°F
Repeatability & Hysteresis	+ 0.1°F
Accuracy	+ 0.02°F

C. Dew Point Measurement System

Inst #	DPT-31 through DPT-40
Mfg	Volumetrics
Model	Dewcell
Range	Calibrated for 50°, 80°, 100°F
Accuracy	+ .00726 psi
Repeatability	+ .002 psi
Stability (for three days)	+ .00006 psi

D. Readout System

Mfg	Data logger	
Model	3000	
accuracy	+ .19°F	+ .0024 psi
repeatability	+ .05°F	+ .0006 psi
stability	+ .05°F	+ .0006 psi

E. Flow Measurement System

Inst #	FI-1
Mfg	Fisher-Porter
Model	087-615-588
Calibration	.4 to 4 scfm of air @ 60 psig & 80°F
Accuracy	$\pm 1\%$ of full scale

II. Formula

$$E_L = 100 \sqrt{2 \left(\frac{E_p}{P-P_V} \right)^2 + 2 \left(\frac{E_T}{T} \right)^2 + 2 \left(\frac{E_V}{P-P_V} \right)^2}$$

where:

E_L = error in leakage rate in % by weight per day

E_p = pressure instrument error $\sqrt{\text{no. inst.}}$

= error in total absolute pressure in psia

E_T = temperature instrument error $\sqrt{\text{no inst.}}$

= error in temperature in °R

E_V = vapor pressure instrument error $\sqrt{\text{no. inst.}}$

= error in water vapor pressure in psia

Assuming:

A. Initial and final temperatures are identical

B. Initial and final pressures are identical

III. Calculation of E_p

$$\begin{aligned} E_p &= \sqrt{\frac{(\text{repeatability})^2 + (\text{resolution})^2 + (\text{stability})^2}{\# \text{ of sensors}}} \\ &= \sqrt{\frac{(.0005)^2 + (.001)^2 + (.002)^2}{2}} \\ &= 1.6202 \times 10^{-3} \text{ psi} \end{aligned}$$

IV. Calculation of E_T

$$E_T = \sqrt{\frac{(\text{repeatability of RTD})^2 + (\text{repeatability of data logger})^2 + (\text{stability of data logger})^2}{\# \text{ sensors}}}$$

$$= \sqrt{\frac{(.1)^2 + (.05)^2 + (.05)^2}{30}}$$

$$= .02236 {}^\circ\text{F}$$

V. Calculation of E_V

$$E_V = \sqrt{\frac{(\text{repeatability of detector})^2 + (\text{stability of detector})^2 + (\text{repeatability of readout})^2 + (\text{stability of readout})^2}{\# \text{ of sensors}}}$$

$$= \sqrt{\frac{(.002)^2 + (.00006)^2 + (.0006)^2 + (.0006)^2}{9}}$$

$$= .00072 \text{ psi}$$

VI. Calculation of E_L

$$E_L = 100 \sqrt{2 \left(\frac{E_p}{P-P_v} \right)^2} = 2 \left(\frac{E_T}{T} \right)^2 + 2 \left(\frac{E_V}{P-P_v} \right)^2$$

$$= 100 \sqrt{2 \left(\frac{.0016202}{74.7} \right)^2 + 2 \left(\frac{.02236}{530} \right)^2 + 2 \left(\frac{.00072}{74.7} \right)^2} = .0068\% \text{ by wt/day}$$

VII. Calculation of E_F

Standard conditions for the flowmeters are defined as 14.73 psia and 60°F.

$$\text{Flowmeter Accuracy} = \frac{1 \times 4.0}{100} = .04 \text{ scfm}$$

$$\text{Flowmeter Error} = E_F = \frac{P_{60}}{T_{80}} \times \frac{T}{P-P_v} \times \frac{60 \times 24 \times 100}{1.05 \times 10^6} \times .04$$

$$= \frac{60 + 14.7}{80 + 460} \times \frac{76 + 460}{75.107 - .143} \times .137 \times .04 = .0054$$

VIII. Calculation of E_L'

$$E_L' = E_L + E_F = .0068 + .0054 = .0122\% \text{ by wt/day}$$