

APPENDIX

NOTICE OF VIOLATION

Comanche County Memorial Hospital
Lawton, Oklahoma

Docket 30-12272/89-01
License 35-17144-01

During an NRC inspection conducted on October 31, 1989, a violation of NRC requirements was identified. In accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C (1989) (Enforcement Policy), the violation is listed below:

10 CFR 35.50(b)(3) requires that the dose calibrator linearity test be performed over the range of its use between the highest dosage that will be administered to a patient and 10 microcuries.

Contrary to the above, as of October 31, 1989, dose calibrator linearity tests were not performed down to the range of 10 microcuries.

This is a Severity Level IV violation. (Supplement VI)

Pursuant to the provisions of 10 CFR 2.201, Comanche County Memorial Hospital is hereby required to submit to this office, within 30 days of the date of the letter transmitting this Notice, a written statement or explanation in reply, including for each violation: (1) the reason for the violation if admitted, (2) the corrective steps which have been taken and the results achieved, (3) the corrective steps which will be taken to avoid further violations, and (4) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending the response time.

Dated at Arlington, Texas,
this 30th day of November, 1989