

030-20721

NRC Form 314  
(11-82)  
10 CFR 30.36(a)(1)(i)(v)  
10 CFR 40.42(a)(1)(i)(v)  
10 CFR 70.76(a)(1)(i)(v)

U.S. NUCLEAR REGULATORY COMMISSION

APPROVED BY OMB  
3196-0020  
EXPIRES 12-31-88

### CERTIFICATE OF DISPOSITION OF MATERIALS

(All items MUST be completed, please print)

41692  
7C  
120743

LICENSEE NAME AND ADDRESS Haber DPM, Emanuel M. 168 Franklin Turnpike Waldwick, NJ 07463	LICENSE NUMBER 29-20687-01
	LICENSE EXPIRATION DATE 04-30-89

THE LICENSEE OR ANY INDIVIDUAL EXECUTING THIS CERTIFICATE ON BEHALF OF THE LICENSEE CERTIFIES THAT: (Check and/or complete the appropriate item(s) below.)

#### A. MATERIALS DATA (Check one and complete as necessary)

1. NO MATERIALS HAVE EVER BEEN POSSESSED OR PROCURED BY THE LICENSEE UNDER THIS LICENSE.

OR

2. ALL MATERIALS PROCURED AND/OR POSSESSED BY THE LICENSEE UNDER THE LICENSE NUMBER CITED ABOVE HAVE BEEN TRANSFERRED ON

DATE 10-17-88	TO Lixi, Inc. 1438 Brook Drive, Downers Grove, IL 60515 (312) 620-4646	WHICH HAS NRC LICENSE NUMBER 12-18215-01
------------------	--	---

OR

3. ALL MATERIALS PROCURED AND/OR POSSESSED BY THE LICENSEE UNDER THE LICENSE NUMBER CITED ABOVE HAVE BEEN TRANSFERRED ON

DATE	TO	WHICH HAS LICENSE NUMBER	ISSUED BY THE STATE OF

AN AGREEMENT STATE PURSUANT TO SECTION 274 OF THE ATOMIC ENERGY ACT OF 1954, AS AMENDED, AND THE ENERGY REORGANIZATION ACT OF 1974.

4. MATERIALS HAVE BEEN DISPOSED OF IN THE FOLLOWING MANNER. (Describe specific disposal procedures--if additional space is needed, use the reverse of this form, or provide attachments.)

Date FEB 09 1989	Duplicate of issued central #110034
Log Feb 92	
By S. Kimberly	
Date Completed FEB 09 1989	

**FEE EXEMPT**  
Termination

8912060208 890117  
REG1 LIC30  
29-20687-01 PDR

#### B. OTHER DATA

1. OUR LICENSE HAS NOT YET EXPIRED. PLEASE TERMINATE IT.

2. WAS A RADIATION SURVEY CONDUCTED TO CONFIRM THE ABSENCE OF LICENSED RADIOACTIVE MATERIALS AND TO DETERMINE WHETHER ANY CONTAMINATION REMAINS ON THE PREMISES COVERED BY THE LICENSE? (Check one)

NO

YES. THE RESULTS (Check one)

ARE ATTACHED. OR

WERE FORWARDED TO NRC ON (Date)

3. THE PERSON TO BE CONTACTED REGARDING THE INFORMATION PROVIDED ON THIS FORM

NAME Emanuel M. Haber, DPM	TELEPHONE NUMBER (201) 652-2432
-------------------------------	------------------------------------

4. MAIL ALL FUTURE CORRESPONDENCE REGARDING THIS LICENSE TO

Waldwick Podiatry Center, P.A./Emanuel M. Haber, DPM

168 Franklin Turnpike, Suite 100, Waldwick, New Jersey 07463-1802

ML10  
1/1

RETURN TO	CERTIFYING OFFICIAL	DATE
DIRECTOR, DIVISION OF FUEL CYCLE AND REACTOR TECHNOLOGY OFFICE OF NUCLEAR REGULATORY AFFAIRS U.S. NUCLEAR REGULATORY COMMISSION WASHINGTON, DC 20555	SIGNATURE Emanuel M. Haber, DPM PRINTED NAME AND TITLE Emanuel M. Haber, DPM	01-17-89

OFFICIAL RECORD COPY ML16

Rec'd 1/19/89

120743



**COPY**  
UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
WASHINGTON, D. C. 20555

JANUARY 03, 1989

CANGIALOSI DPM, CHARLES P.  
HABER DPM, EMANUEL M.  
168 FRANKLIN TURNPIKE  
WALDWICK, NJ 07463

License No. 29-20687-01  
Expiration Date: 04/30/89  
Program Code: 02201

Gentlemen:

**SUBJECT: NOTICE OF EXPIRATION**

Your NRC license specified above will expire on the date shown. If you wish to continue your program, you must submit an application for license renewal in accordance with the enclosed instructions.

If your application for license renewal is filed at least 30 days before the expiration date of your license and the application is accompanied by the appropriate fee for license renewal (see 10 CFR 170.31), your license will remain in effect until final action is taken on your application. However, if your application is filed less than 30 days before the expiration date and it cannot be processed before that date, you would be without a valid license when the license expires.

If you do not wish to renew your license, you must dispose of or transfer all licensed radioactive material in your possession in an authorized manner (see 10 CFR 30.36), complete the enclosed Form NRC-314, "Certificate of Disposition of Materials," and return it before the expiration date of your license with a request that your license be terminated. If you cannot dispose of or transfer all licensed radioactive material in your possession before the expiration date, you must request a license renewal for storage only of the radioactive material to avoid violations of possessing licensable material without a valid license.

This notice of your license expiration is sent for your convenience and it should not be interpreted that similar notices will be sent in the future. The responsibility for timely submission of license renewal remains with the licensee.

Nuclear Materials Section  
Division of Engineering and  
Technical Programs

Enclosures:

1. Renewal Instructions
2. Form 313
3. Form 314
4. 10 CFR Parts 20, 30, 35, and 170
5. Reg. Guide 10.8

WALDWICK PODIATRY CENTER, P.A.

168 FRANKLIN TPKE  
WALDWICK, NEW JERSEY 07463  
(201) 652-2432

EMANUEL M. HABER, DPM\*

RECONSTRUCTIVE SURGERY OF THE FOOT & ANKLE  
\*FELLOW, AMERICAN SOCIETY OF PODIATRIC MEDICINE

CERTIFICATION OF RADIATION SURVEY

LOCATION: The Waldwick Podiatry Center, P.A.  
168 Franklin Turnpike, Suite 100  
Waldwick, New Jersey 07463-1802  
(201) 652-2432

LICENSE NUMBER: 29-20687-01

DATE PERFORMED: January 16, 1989

PERFORMED BY: Emanuel M. Haber, DPM

UNIT UTILIZED: Minimonitor II

MODEL #05-571

RATE: mR/h

SERIAL #14132

MANUFACTURER: Nuclear Associates, Division of Victoreen, Incorporated  
RMI Florida Incorporated, Florida 32935, (305) 259-6862

AREA TESTED	DURATION OF TEST	SOURCE SETTING	SOURCE READING
Wall.....	1 hour.....	x 1.....	< 1/2 mR/hour
Counter.....	1 hour.....	x 1.....	"
Drawer.....	1 hour.....	x 1.....	"
Floor.....	1 hour.....	x 1.....	"

It is determined that the level of total remaining contamination on these premises was less than 1 mR per hour.

Emanuel Haber DPM  
Emanuel M. Haber, DPM  
Fellow, American Society of Podiatric Medicine

EMH:baar





COPY

CERTIFICATE OF LEAK TEST

CUSTOMER:

WALD WICK PODIATRY CENTER

SERIAL NO. OF SOURCE 7482

ISOTOPE I<sup>125</sup> ACTIVITY 461mCi

REMARKS \_\_\_\_\_

MANUFACTURER AECL

CALIBRATION TEST DATE 9/29/87

This is to certify that the leak test on the above source was performed by AECL on 9/21/87  
(Testing Facility) (Date)

and validated by D. BONHOWER of that facility.

It was determined that the level of removable contamination was below 0.005 microcurie.

The original certificate is on file at Lixi, Inc.

NOTE: Removable contamination equal to or greater than 0.005 microcurie constitutes a leaking source.

DATE: 9/23/87

SIGNED: John B. Pindler

/kp  
6/6/83

COPY

CERTIFICATE OF LEAK TEST

CUSTOMER:

WALDNICK PODIATRY CENTER PA

SERIAL NO. OF SOURCE 5829-8

ISOTOPE I<sup>125</sup> ACTIVITY 400 mCi

REMARKS \_\_\_\_\_

MANUFACTURER AELL

CALIBRATION TEST DATE 5/11/86

This is to certify that the leak test on the above source was performed by AELL on 5/6/86  
(Testing Facility) (Date)

and validated by T. SIMPSON of that facility.

It was determined that the level of removable contamination was below 0.005 microcurie.

The original certificate is on file at Lixi, Inc.

NOTE: Removable contamination equal to or greater than 0.005 microcurie constitutes a leaking source.

DATE: 5/14/86

SIGNED: John G. Powell

/kp  
6/6/83

COPY

STAN A. HUBER CONSULTANTS, INC.

-- 200 N. CEDAR ROAD -- NEW LENOX IL 60451 -- PHONE (815) 485-6161 --

===== SEaled GAMMA/BETA SOURCE - LEAK TEST CERTIFICATE =====

LEAK TEST REPORT DATE: 05-28-86  
FACILITY: WALDWICK PODIATRY CENTER  
WALDWICK NJ  
FACILITY RADIOACTIVE MATERIALS LICENSE NO.: 29-20687-01  
 NRC -- STATE

SOURCE IDENTIFICATION

RADIONUCLIDE: I-125                      ACTIVITY: 400 MCI  
CALIBRATION DATE: 05-11-86            MANUFACTURER: LIXI  
MODEL NO.:                                SERIAL NO.: 5829-B

ASSAY RESULTS:

	GROSS COUNTS/MIN	NET COUNTS/MIN	UCI
WET WIPE	67	0	< 9.1E-07
DRY WIPE	61	0	< 9.1E-07

BACKGROUND @ TIME OF ASSAY 69 CPM

ANALYSIS OF RESULTS:

SOURCE IS NOT LEAKING AT THIS TIME.  
REMOVABLE CONTAMINATION IS LESS THAN  $5 \times 10E-3$  UCI.

NEXT LEAK TEST DUE: NOVEMBER 1986

LEAK TEST PERFORMED BY: C. P. CANGIALOSI

DATE: 05-20-86

ANALYSIS PERFORMED BY: STAN A. HUBER

DATE: 05-27-86



LIXI

# COPY

## CERTIFICATE OF LEAK TEST

CUSTOMER:

DR CHARLES P. CANUCCI

SERIAL NO. OF SOURCE 5156-B

ISOTOPE I<sup>125</sup> ACTIVITY 432 mCi

REMARKS \_\_\_\_\_

MANUFACTURER AECL

CALIBRATION TEST DATE 9/9/86

This is to certify that the leak test on the above source was performed by AECL on 8/28/86  
(Testing Facility) (Date)

and validated by T. SIMPSON of that facility.

It was determined that the level of removable contamination was below 0.005 microcurie.

The original certificate is on file at Lixi, Inc.

NOTE: Removable contamination equal to or greater than 0.005 microcurie constitutes a leaking source.

DATE: 9/9/86

SIGNED: John B. ...

/kp  
6/6/83



COPY

CERTIFICATE OF LEAK TEST

CUSTOMER:

DR CHARLES CANGIALOSI

SERIAL NO. OF SOURCE 4158

ISOTOPE I<sup>125</sup> ACTIVITY 400.mCi

REMARKS \_\_\_\_\_

MANUFACTURER AECL

CALIBRATION TEST DATE 4/29/85

This is to certify that the leak test on the above source was performed by AECL on 4/29/85  
(Testing Facility) (Date)

and validated by T. SIMPSON of that facility.

It was determined that the level of removable contamination was below 0.005 microcurie.

The original certificate is on file at Lixi, Inc.

NOTE: Removable contamination equal to or greater than 0.005 microcurie constitutes a leaking source.

DATE: 5/1/85

SIGNED: John D. Pindick

/kp  
6/6/83



LIXI

COPY

CERTIFICATE OF LEAK TEST

CUSTOMER:

DR. CHARLES PHILIP  
CANGIPLDSI  
\_\_\_\_\_  
\_\_\_\_\_

SERIAL NO. OF SOURCE 5235  
ISOTOPE I<sup>125</sup> ACTIVITY 412 mCi  
REMARKS \_\_\_\_\_  
MANUFACTURER AECL  
CALIBRATION TEST DATE 12/3/85

This is to certify that the leak test on the above source was performed by AECL on 11/26/85 (Testing Facility) (Date) and validated by T. SIMPSON of that facility. It was determined that the level of removable contamination was below 0.005 microcurie.

The original certificate is on file at Lixi, Inc.

NOTE: Removable contamination equal to or greater than 0.005 microcurie constitutes a leaking source.

DATE: 11/27/85

SIGNED: [Signature]

/kp  
6/6/83

COPY

CERTIFICATE OF LEAK TEST

CUSTOMER:

Dr. Charles Cangiolosi

SERIAL NO. OF SOURCE 3060

ISOTOPE I<sup>125</sup> ACTIVITY 412 mCi

REMARKS —

MANUFACTURER AECL

CALIBRATION TEST DATE 5/11/84

This is to certify that the leak test on the above source was performed by AECL on 5/3/84  
(Testing Facility) (Date)

and validated by Bruce Raby of that facility.

It was determined that the level of removable contamination was below 0.005 microcurie.

The original certificate is on file at Lixi, Inc.

NOTE: Removable contamination equal to or greater than 0.005 microcurie constitutes a leaking source.

DATE: 5/7/84

SIGNED: Bruce Van Pelt

/kp  
6/6/83

COPY

CERTIFICATE OF LEAK TEST

CUSTOMER:

Dr. Charles Congia Lasi  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SERIAL NO. OF SOURCE 3570  
ISOTOPE I<sup>125</sup> ACTIVITY 437mCi  
REMARKS \_\_\_\_\_  
MANUFACTURER AELL  
CALIBRATION TEST DATE 10/3/84

This is to certify that the leak test on the above source was performed by AELL on 10/24/84  
(Testing Facility) (Date)

and validated by J. W. Simpson of that facility.

It was determined that the level of removable contamination was below 0.005 microcurie.

The original certificate is on file at Lixi, Inc.

NOTE: Removable contamination equal to or greater than 0.005 microcurie constitutes a leaking source.

DATE: 10/26/84

SIGNED: 

/kp  
6/6/83

OFFICIAL RECORD COPY ML10

120743



(FOR LFMS USE)  
INFORMATION FROM LTS

BETWEEN:

License Fee Management Branch, ARM  
and  
Regional Licensing Sections

Program Code: 02201  
Status Code: 0  
Fee Category: 7C  
Exp. Date: 19890430  
Fee Comments: \_\_\_\_\_

LICENSE FEE TRANSMITTAL

A. REGION 1

1. APPLICATION ATTACHED

Applicant/Licensee: CANGIALOSI DPM, CHARLES P.  
Received Date: 890119  
Docket No.: 3020721  
Control No.: 120743  
License No.: 29-20687-01  
Action Type: Termination

2. FEE ATTACHED

Amount: \_\_\_\_\_  
Check No.:    /   /   

3. COMMENTS

Signed M. Moriarty  
Date 01-23-89

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is achieved)

1. Fee Category and Amount: 7C

2. Correct Fee Paid. Application may be processed for  
Amendment   
Renewal \_\_\_\_\_  
License \_\_\_\_\_

**FEE EXEMPT**  
Termination

3. OTHER \_\_\_\_\_

Signed J. Kimberly  
Date FEB 09 1989