

NUCLEAR  
RESEARCH  
CORPORATION

1105 Industrial Highway  
Southampton, Pa. 18966  
Phone: 215 - 357 - 5015  
215 - 464 - 4434

April 7, 1981

Nuclear Regulatory Commission  
Region I  
631 Park Avenue  
King of Prussia, PA 19406

Attention: John D. Kinneman

Subject: Inspection 81-01


Gentlemen:

We wish to acknowledge your letter of 27 March 1981 in regard to Wipe Testing the Nuclear Research Corporation UDM-1A Cesium Calibrator Source.

Please be advised that corrective action was taken on 4 March 1981 as well as assurance that Wipe Tests on sealed sources containing by-product material will be performed at intervals not to exceed six months.

Sincerely,

NUCLEAR RESEARCH CORPORATION

  
E. M. Polluck  
President

mg

8106040456