

# Augustana

HOSPITAL AND  
HEALTH CARE CENTER

411 WEST DICKENS AVENUE  
CHICAGO, ILLINOIS 60614  
312 975 5000

May 22, 1981

Mr. D. G. Wiedeman, Acting Chief  
Materials Radiation Protection  
Section I  
c/o U. S. Nuclear Regulatory Commission  
Region III  
799 Roosevelt Road  
Glen Ellyn, Illinois 60137


Dear Mr. Wiedeman:

This letter is written in regards to your letter of May 13, 1981 detailing a Notice of Violation for our License 12-00710-02. The specific citation was that of our Nuclear Medicine Department personnel not surveying upon receipt packages containing byproduct material.

Effective May 4, 1981, we took corrective action which should avoid future noncompliance with this item. Henceforth from the above date, we have and will continue to keep a logbook detailing material received, date recieved, technologist accepting material and results of leakage testing. This information will be readily available for future inspections.

I hereby do affirm the above is true, May 22, 1981.

Sincerely,



John H. Reising  
Vice President  
Medical Affairs

JHR:mcm

cc: J. Desamero  
T. Patel  
K. Roman

8106190287

JUN 1 1981