


Henrotin Hospital

111 West Oak Street
Chicago, Illinois 60610
Phone 312.440.7700



May 1, 1981

D. G. Wiedeman
Acting Chief Materials Radiation Protection
Section 1
United States Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137

RE: License No. 12-13711-01

Dear Mr. Wiedeman:

In reference to your letter dated April 27, 1981, please be advised of those items of non-compliance have been and are being dealt with in the following manner:

- Item #1. The linearity and accuracy tests of the dose calibrator are being performed and our radiation physicist, Mr. Ron Edwards, will forward records of past tests to your department.
- Item #2. Survey of all incoming packages of radiopharmaceuticals is now being reported and the records are being kept with the dose surveys.
- Item #3. The daily survey procedures have been expanded to include those areas involving all elution, preparation, and injection of radiopharmaceuticals.

In regard to the final paragraph of your letter, please be advised that the above items represent the corrective action taken. A quarterly report to the Radiation Protection And Isotope Committee will avoid any further non-compliance in these areas in the future. The date of full compliance is the following 5/1/81.

Should there be any further questions in regard to the action taken on your recommendations please advise me in writing. Thanking you for your consideration I remain.

Yours truly,

Subscribed and Sworn to before me
this 1 day of May 1981. Notary
Public Virginia M. T. [Signature]

Lawrence Schlachter
Lawrence Schlachter, M.D.
Chairman-Department of Radiology

My Commission Expires June 8, 1981

LS:mc

81 45260186

MAY 4 1981