## NOTICE OF VIOLATION

Globe X-Ray Services, Inc.

Docket No. 30-8917 License No. 35-15194-01

As a result of the inspection conducted on May 6, 1981, and in accordance with the attached Interim Enforcement Policy, 45 FR 66754 (October 7, 1980), the following violations were identified:

 10 CFR 34.33(c) requires that pocket dosimeters shall be checked for correct response to radiation to insure dosimeters read within +/-30% of the true radiation exposure.

Contrary to this requirement, pocket dosimeters had not been checked for correct response to radiation during the period March 1, 1979 to May 6, 1981.

This is a Severity Level IV violation (Supplement VII.D.2).

2. 10 CFR 71.5 requires that shipments of licensed materials be made in compliance with applicable regulations of the Department of Transportation. 49 CFR 173.398(a) requires, in part, that each shipper of special form radioactive material shall maintain on file for at least one year after the last shipment a complete certification and supporting analysis of special form material shipped by the licensee.

Contrary to this requirement, sealed source performance test records as required by 49 CFR 173.398(a) were not available for sources shipped during the period March 1, 1979 to May 7, 1981.

This is a Severity Level V violation (Supplement VII.E).

Pursuant to the provisions of 10 CFR 2.201, Globe X-Ray Services, Inc., is hereby required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) the corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Under the authority of Section 182 of the Atomic Energy Act of 1954, as amended, this response shall be submitted under oath or affirmation. Consideration may be given to extending your response time for good cause shown.

Dated	May 21, 1981	of west for
		Glen D. Brown, Chief
		Technical Inspection Branch