



CARDINAL CUSHING GENERAL HOSPITAL

235 NORTH PEARL STREET, BROCKTON, MASSACHUSETTS 02401 (617) 588-4000

November 3, 1980

Mr. John D. Kinneman, Chief
Materials Radiological Protection
Section
United States Nuclear Regulatory
Commission
Region 1
631 Park Avenue
King of Prussia, Pennsylvania 19406


Docket Nos. 30-01935
70-02529

Dear Mr. Kinneman:

Thank you for your letter of September 26. We have followed your suggestion to investigate the possibility of using other styles of syringe shields and have found an all leaded glass syringe shield manufactured by Nuclear Pacific which allows us to adequately see the accurate dose as it is being withdrawn from the container. Consequently, we are simply going to obtain an adequate number of these syringe shields and use them in accordance with our present license. This should put us in full compliance with the violation number 1 in our inspection, number 80-01.

Thank you very much for your suggestion and consideration.

Sincerely,


Robert J. Jepsen, Jr.
Administrator

jek

810 1290626