


- B. 2. Our present procedure states that ¹³⁷Cesium shall be assayed using a reproducible geometry before each daily use of the instrument. The following statement will be added to the procedure; "As soon as the assay is performed the results will be placed into a ledger in which the date is clearly entered."
We feel that the non-compliances which occurred in the past were assays that were performed but were not logged. This procedure will be in effect immediately.
3. Our procedure manual will be amended to read: "Linearity tests will be performed in each calendar quarter" This will become effective immediately.
- C. During a subsequent re-examination of our records, we located records indicating that wipe tests were performed on the following dates: July 6, 1978, December 27, 1978, April 9, 1979, October 26, 1979, March 27, 1980, September 23, 1980 and January 29, 1981.

If there any questions concerning this matter, please let us know.

Sincerely,


Sister M. Elizabeth Corry, OSF
Executive Director

SMEC:vf

cc: Sister Norma Matthews, OSF
Assistant Administrator

Deborah Hitchcock
Supervisor, Nuclear Medicine

Ralph Caccese, M.D.
Department of Radiology