



**RIVERSIDE HOSPITAL**

1600 SUPERIOR STREET

**TOLEDO, OHIO 43604**

May 8, 1981

TELEPHONE  
419-729-5151

NUCLEAR REGULATORY COMMISSION  
Region III  
799 Roosevelt Road  
Glen Ellyn, Illinois 60137

Attention: Mr. D. J. Sreniawski

Dear Sir:

Attached herewith is a statement of the corrective actions taken by the Nuclear Medicine Department of Riverside Hospital in response to your letter of April 30, 1981.

These documents are submitted under oath as indicated below.

Sincerely,

*Harold Crary*

Harold Crary  
Director  
Professional Services

HC/ci

att. (1)

cc: Nuclear Medicine Dept.  
File

*May 8, 1981*  
*Anna M. Pitzen*  
*Notary Public*

8106010 455

ANNA M. PITZEN  
Notary Public, Lucas County, Ohio  
My commission expires Feb. 20, 1986

MAY 11 1981

Violation 1. Linearity Tests

1. Corrective action taken and results achieved - A linearity test was done on April 27 by departmental personnel and recorded. These tests will be done on a quarterly basis with the results kept on file for future inspection.
2. Corrective action to be taken to avoid future noncompliance - The linearity tests will be checked periodically by an outside radiation health physicist to assure accurate testing is being done.
3. The date when full compliance will be achieved - Full compliance has been achieved as of April 27, 1981 when a second linearity test for 1981 was performed. A third test will be done in July, 1981.

Violation 2. Survey of Packages

1. A form has been derived from the NRC Regulatory Guide and adapted to provide for recording package surveys and wipe testing. All incoming packages are now surveyed on arrival and the results recorded.
2. Corrective action to be taken to avoid future noncompliance - These records will be checked periodically by an outside radiation health physicist to assure that proper records are being maintained.
3. The date when full compliance will be achieved - Full compliance was achieved on April 20, 1981.

Violation 3. Area Survey and Wipe Testing

1. Corrective action taken and the results achieved - Isotope preparation areas and injection areas are now surveyed on a daily basis and wipe tested weekly. This is in addition to a monthly survey and wipe test performed by an outside radiation health physicist. The results of these surveys and wipe tests are recorded and on file for future inspection.
2. Corrective action to be taken to avoid future noncompliance - These records will be checked periodically by an outside radiation health physicist to assure that proper records are being maintained.
3. The date when full compliance will be achieved - Full compliance was achieved on April 20, 1981.

May 8, 1981

*Anna M. Pitzen, Notary Public*  
ANNA M. PITZEN

Notary Public, Lucas County, Ohio

My commission expires Feb. 20, 1985