

## Johnson County Memorial Hospital

(317) 736-3300

FRANKLIN, INDIANA 46131

P. O. BOX 607 1125 WEST JEFFERSON

March 28, 1981

United States Nuclear Regulatory Commission D. Sreniawski, Acting Chief Region III 799 Roosevelt Road Glen Ellyn, Illinois 60137

## Gentlemen:

This letter is in response to findings of non-compliance indicated by Mr. W. P. Reichhold while performing a routine inspection on February 4, 1981 of Johnson County Memorial Hospital, License No. 13-14817-01.

ITEM NO. 1: In accordance with License Condition No. 15, the Nuclear Medicine Committee shall meet on a quarterly basis.

ITEM NO. 2: As indicated by our license statements, a wipe test of each package containing radioactive materials will be performed and recorded.

ITEM NO. 3: In compliance with our license requirements, the "hot laboratory area" will be surveyed with a G-M meter and wipe tests performed on a daily basis.

ITEM NO. 4: Daily constancy checks of the dose calibrator with Cs-137 and Co-57 will be performed as indicated by our license statements.

ITEM NO. 5: Physical inventory and wipe tests of the sealed calibration sources will be performed on a quarterly basis as indicated by our license requirements.

These procedures shall be accomplished on a regular basis as indicated and shall begin as follows.

ITEM NO. 1: Nuclear Medicine Committee meetings shall occur on or about the first week of January, April, June and October. The next scheduled meeting will be the first week of April. Page 2 March 28, 1981 United State Nuclear Regulatory Commission

ITEMS NO. 2, 3, AND 4: Will be performed at the suggested frequency beginning immediately.

ITEM NO. 5: Will be performed on a quarterly basis on or about the first week of December, March, May and September to allow evaluation and review of any complications prior to scheduled Nuclear Medicine Committee meetings. The wipe tests and inventory have been performed and the next scheduled performance is May, 1981.

To prevent oversight in the future, a calendar is being prepared to allow quick review of the necessary surveys and meetings.

Sincerely,

Arthur Gross,

Assistant Administrator