Diagnostic Pad diogy Nuclear Scans Card at Testing Ultrasound Pulmonary Function EEG—EMG

## BIOMEDICAL DIAGNOSTIC SERVICES

3270 W. BIG BEAVER RD. • SUITE 115 • TROY, MICHIGAN 48084 (313) 649-4449

April 13, 1981

Melvin A. Lester. M.D.

And Response:

Time Its Int w/

Hirmatin ettale

D. J. Sreniawski
Acting Chief
Materials Radiation Protection, Section 2
U. S. Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, IL 60137

Re: Response - Notice of Violation

Dear Mr. Sreniawski:

The following is in response to your letter dated 3/26/81 which outlined the results of a routine safety inspection conducted in our Nuclear Medicine Section under Materials License No. 21-17122-01. Our response to this Notice of Violation will follow the outline of Appendix A of your letter persuant to 10 CFR 2.201 as follows:

- 1) Daily constancy checks of our dose calibrator were initiated on 3/10/81 and our results have noted a consistancy of precision of ± 5%. Dose calibrator linearity was tested on 3/12/81 and linearity was within ± 5% of true linear response. We confirm that the daily constancy check is now a regular daily routine of our Nuclear Medicine Section and our dose calibrator will be tested for linearity on a quarterly schedule.
- 2) A semiannual test for leakage of our Cesium-137 calibration source was initiated on 2/25/81 and results indicated no leakage at that time (3 x  $10^{-6}$  uCi). We confirm that this test for leakage will be performed on a semiannual basis.
- 3) Daily GM radiation area surveys were initiated on 3/11/81 and weekly wipe tests for contamination on 3/13/81. Results of both thus far have indicated no contamination. We confirm that radiation area surveys will be conducted daily and wipe tests performed on a weekly basis.

If you have any questions, please contact me.

Sincerely,

Shirley V. Bosley
Assistant Administrator

I affirm the above statements are true.

Shirley Bosley, Assistant Administrator

SVB/ldj enc.

8106010491