

Appendix A

NOTICE OF VIOLATION

St. Joseph Hospital

License No. 12-03832-02

As a result of the inspection conducted on April 22, 1981, and in accordance with the Interim Enforcement Policy, 45 FR 66754 (October 7, 1980), the following violations were identified:

1. License Condition No. 18 requires that licensed material be possessed and used in accordance with statements, representations and procedures contained in application dated June 1, 1978, and referenced letters. This application states that: (1) area radiation (GM) surveys will be performed daily, (2) wipe tests for removable contamination will be performed on Mondays, Wednesdays, and Fridays, and; (3) records of all surveys will be maintained.

Contrary to the above, area surveys and wipe tests were performed only once during November 1978, and January 1979, three times during August 1979, and seven times during October 1979; no surveys were performed between January 26 and July 24, 1979, April 1 through 22, 1979, and December 11 through 29, 1980; and no records were available for the months of May and December 1978, or September 1979, and several other occasions.

This is a Severity Level V violation (Supplement VII).

2. License Condition No. 18 requires that licensed material be possessed and used in accordance with statements, representations and procedures contained in application dated June 1, 1978, and referenced letters. This application states that your Radiation Safety Committee will meet at least quarterly.

Contrary to the above requirement, your Radiation Safety Committee did not meet during the third or fourth quarter 1979, or third quarter 1980.

This is a Severity Level V violation (Supplement VII).

3. License Condition No. 18 requires that licensed material be possessed and used in accordance with statements, representations, and procedures contained in application dated June 1, 1978, and referenced letters. This application states that linearity tests on your dose calibrator will be performed semiannually.

8106030054


Contrary to the above requirement, linearity tests were performed on March 1980, and February 1981, a period in excess of the semiannual frequency required.

This is a Severity Level V violation (Supplement VII).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within twenty-five days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Under the authority of Section 182 of the Atomic Energy Act of 1954, as amended, this response shall be submitted under oath or affirmation.

5-8-81

Dated



D. G. Wiedeman, Acting Chief
Materials Radiation Protection
Section 1