

CAPITOL HILL HOSPITAL

" WE CARE "



April 29, 1981

Mr. John D. Kinneman, Chief
U.S. Nuclear Regulatory Commission
Materials Radiological Protection
Section
631 Park Avenue
King of Prussia, Pa. 19406

Dear Mr. Kinneman:

In a letter dated October 29, 1980, you reported the results of Inspection 80-01 of the activities conducted under NRC License No. 08-08327-02. The last sentence in that letter states that "no reply to this letter is required." We were recently notified by phone that a reply was expected.

The following statements are presented in response to the two infractions specified in Appendix "A" of the October 29, 1980 letter.

1. The linearity of the dose calibrator is now being performed on a quarterly basis using a Tc-99m source. The latest test was performed on December 17, 1980.
2. All of the sealed calibration sources possessed pursuant to 10CFR-35.14(d) are now being tested for leakage every six months. The latest series of such tests were performed on January 13, 1981.

Sincerely,

Luis M. Lopez M.D.
Luis M. Lopez, M.D.

8106090421