



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
631 PARK AVENUE
KING OF PRUSSIA, PENNSYLVANIA 19406

Docket Nos. 30-2464
30-00347

JUN 1981



Hospital Center at Orange
ATTN: Mr. Arthur T. Dunn
President
188 South Essex Avenue
Orange, New Jersey 07051

Gentlemen:

Subject: Combined Inspection 30-2464/81-01 and 30-00347/81-01


This refers to your letter dated April 1, 1981, in response to our letter dated March 10, 1981.

Thank you for informing us of the corrective and preventive actions documented in your letter. These actions will be examined during a future inspection of your licensed program.

From the telephone conversation between Miss J. McGinness of this office and Dr. Sinclair, of your staff on April 23, 1981, it is our understanding that the Hospital Center's dosimetry system will be calibrated at a Regional Calibration Center prior to your next full calibration of your teletherapy unit. In the interim, you are planning to use a system from another hospital which has been calibrated as required by 10 CFR 35.23(a) to perform the monthly spotchecks required by 10 CFR 35.22(b). While this office is aware of the delays associated with having instruments calibrated at the Regional Calibration Laboratory, you are advised that you will continue to be in noncompliance until your corrective actions are complete.

Your cooperation with us is appreciated.

Sincerely,


John D. Kinneman, Chief
Materials Radiological Protection
Section

1E07
S/11

8106090614