Orange Memorial Hospital Unit New Jersey Orthopaedic Hospital Unit



188 South Essex Avenue / Orange, New Jersey 07051 201-266-2200

April 1, 1981

Mr. John D. Kinneman Chief, Materials Radiological Protection Section Technical Inspection Branch U.S. Nuclear Regulatory Commission Region 1 631 Park Avenue King of Prussia, Pennsylvania 19406

Subject: Combined Inspection 30-02464/81-01; 30-00347/8101

Dear Mr. Kinneman:

In reply to your letter dated March 10, 1981, Docket Number 30-0347, License Number 29-03038-02, the following information was supplied by Ian P. Sinclair, Ph.D., radiation physicist to the Hospital Center at Orange.

- A. Checks of timer accuracy and field uniformity were carried out on March 5, 1981. The determination of timer accuracy was carried out according to ANSI N449 section 4:14.1. These determinations will be repeated at intervals of less than one year.
- B. The monthly spot check of the Cobalt 60 output was not carried out in January 1981 due to the repair of the Capintec meter. The spot check for March was carried out on March 5, 1981. In the future if the Hospital Center's meter is not functioning, the spot check will be carried out using a meter borrowed from St. Joseph's Memorial Hospital and Medical Center, Paterson, New Jersey. This meter has a calibration performed by the North East Regional Calibration Center at Memorial Sloan Kettering.
- C. The Hospital Center's dosimetry system was calibrated by comparison with a meter which had a recent calibration at the Regional Center; however, arrangements have been made to have a calibration carried out on the Hospital Center's c n meter at the North East Regional Calibration Center at Memorial Sloan Kettering.

8106090620

Mr. John D. Kinneman

We hope the above information is satisfactory.

Sincerely,

Arthur T. Dunn President

ATD:bay

.

6 8

cc: James M. Monaghan, M.D. Directory of Radiology Ian P. Sinclair, Ph.D. Radiation Physicist