## APPENDIX A

## NOTICE OF VIOLATION

Department of Health and Human Services U.S. Public Health Service Hospital Baltimore, Maryland 21211 Docket No. 30-11392 License No. 19-03520-05

As a result of the inspection conducted on March 4, 1981, and in accordance with the Interim Enforcement Policy, 45 FR 66754 (October 7, 1980), the following violation was identified.

Condition 21 of your license requires that licensed material be possessed and used in accordance with the statements, representations and procedures contained in your application dated April 30, 1976 and letters dated October 18, 1976 and August 11, 1980.

Supplemental Sheet 1 of this application contains in the fourth paragraph on the fourth page of the section titled, "Radiation Source Procurement," requires that when radioactive material is shipped, the Radiation Safety Officer must be notified and a copy of the recipient's license must be furnished in advance.

Contrary to this requirement, a Radiation Machinery Corporation EC1-600A6 extracorporeal blood irradiator containing approximately 14.5 curies stronium-90 was shipped in February 1981 from U.S. PHS Hospital without knowledge of the Radiation Safety Officer and a copy of the recipient's license was not furnished in advance.

This is a Severity Level V Violation (Supplement VII).

Pursuant to the provisions of 10 CFR 2.201, the U.S. Public Health Service Hospital, Baltimore, Maryland, is hereby required to submit to this office within twenty-five days of the date of this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Under the authority of Section 182 of the Atomic Energy Act of 1954, as amended, this response shall be submitted under oath or affirmation.

Dated	2 APR 1981	Joe Timeman
		John D. Kinneman, Chief Materials Radiological Protection Section