

Appendix A

NOTICE OF VIOLATION

Silver Cross Hospital

License No. 12-01127-02

As a result of the inspection conducted on April 28, 1981, and in accordance with the Interim Enforcement Policy, 45 FR 66754 (October 7, 1980), the following violation was identified:

10 CFR 30.3 states that no person shall possess or use byproduct material except as authorized by a specific or general license.


Contrary to this requirement, as of April 28, 1981, you were in possession of a 50 millicurie strontium-90 eye applicator which is not authorized by your license.

This is a Severity Level V violation (Supplement VII).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within twenty-five days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Under the authority of Section 182 of the Atomic Energy Act of 1954, as amended, this response shall be submitted under oath or affirmation.

5-11-81

Dated



D. G. Wiedeman, Acting Chief  
Materials Radiation Protection  
Section 1

8106120421