

APPENDIX A

NOTICE OF VIOLATION

Branch Radiographic Laboratories, Inc.
Cranford, New Jersey 07016
License No. 29-03405-02

Docket No. 30-05291


As a result of the inspection conducted on March 25, 1981, and in accordance with the Interim Enforcement Policy, 45 FR 66754 (October 7, 1980), the following violation was identified:

10 CFR 34.24 requires that each survey instrument used to conduct physical radiation surveys be calibrated at intervals not to exceed three months and after each instrument servicing.

Contrary to this requirement, during the period from January 4, 1981 to March 25, 1981, you conducted physical radiation surveys with a survey instrument which was calibrated more than three months prior to the date of use.

Pursuant to the provisions of 10 CFR 2.201, Branch Radiographic Labs, Inc. is hereby required to submit to this office within twenty-five days of the date of this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Under the authority of Section 182 of the Atomic Energy Act of 1954, as amended, this response shall be submitted under oath or affirmation.

Dated 16 APR 1981



John D. Kinneman, Chief
Materials Radiological Protection
Section

2106180145