

SHIELDS L DALTROFF VICE PRESIDENT ELECTRIC PRODUCTION

PHILADELPHIA ELECTRIC COMPANY

PHILADELPHIA, PA. 19101

(215) 841-5001

April 27, 1981

Docket Nos. 50-277 50-278

Inspection No. 50-277/81-01 50-278/81-01

Mr. Gary Snyder, Chief
Emergency Preparedness and Program
Support Branch
Region 1
U.S. Nuclear Regulatory Commission
631 Park Avenue
King of Prussia, PA 19406

Dear Mr. Snyder:

Your letter of April 3, 1981, forwarded combined Inspection Report Nos. 50-277/81-01 and 50-278/81-01. Appendix A addresses one item which does not appear to be in full compliance with Nuclear Regulatory Commission requirements. This item is categorized as a Severity Level V Violation in accordance with your Interim Enforcement Policy, 45FR 66754 (October 7, 1980), and is restated below with our response.

NONCOMPLIANCE

Section 6.8.1 of the Technical Specifications requires that procedures shall be established, implemented, and maintained that meet the requirements of Appendix "A" of USNRC Regulatory Guide 1.33. Section 6.8.2 of the Technical Specifications requires that each procedure required by Section 6.8.1 of the Technical Specifications be eviewed by the PORC and approved by the Station Superinter ont prior to implementation... Regulatory Guide 1.33 requires procedures for both the operation and calibration of instrumentation used for radioactive effluent analyses.

Contrary to the above, as of January 15, 1981, the written procedure for the calibration of the computer based gamma spectroscopy system used for effluent analyses, was not reviewed

8106300392

POOR ORIGINAL

and approved as required by Sections 6.8.1 and 6.8.2 of the Technical Specifications.

This is a Severity Level V Violation (Supplement I).

RESPONSE:

The procedures utilized for calibration of the computer based gamma spectroscopy system used for radioactive effluent analysis have been reviewed by PORC and approved without comment.

Very truly yours,

COMMONWEALTH OF PENNSYLVANIA :

ss.

COUNTY OF PHILADELPHIA

S. L. Daltroff, being first duly sworn, deposes and says:

That he is Vice President of Philadelphia Electric Company; that he has read the foregoing response to Inspection Report Nos. 50-277/81-01 and 50-278/81-01 and knows the contents thereof; and that the statements and matters set forth therein are true and correct to the best of his knowledge, information and belief.

Subscribed and sworn to

before me this 27 day

of

Totaly Public ELIZABETH H. BOYER

My Commission Expires Jan. 30, 1982

POOR ORIGINAL