

NOTICE OF VIOLATION

Intercity Radiological Labs, Inc.

Docket: 30-07507
License: 25-14896-01

As a result of the inspection conducted on June 3, 1981, and in accordance with the Interim Enforcement Policy, 45 FR 66754 (October 7, 1980), the following violations were identified:

1. 10 CFR 35.22(a) requires, in part, that all teletherapy units used for treating humans shall have spot check measurements as outlined in 10 CFR 35.22(b) performed on each teletherapy unit at intervals not to exceed one month.

Contrary to this requirement, monthly spot check measurements had not been performed at any time during the period August 8, 1977 to June 3, 1981.

This is a Severity Level IV violation (Supplement VII.D.1).

2. 10 CFR 35.25 requires, in part, that all records of full calibration of teletherapy units shall be maintained for inspection by the NRC.

Contrary to this requirement, full calibration records of the teletherapy unit were not available for inspection by a NRC representative.

This is a Severity Level VI violation (Supplement VII.F).

Pursuant to the provisions of 10 CFR 2.201, Intercity Radiological Labs, Inc., is hereby required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) the corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Under the authority of Section 182 of the Atomic Energy Act of 1954, as amended, this response shall be submitted under oath or affirmation. Consideration may be given to extending your response time for good cause shown.

Dated

6/23/81

Glen D. Brown

Glen D. Brown, Chief
Technical Inspection Branch

8106300459