

HADH

HERMANN AREA
DISTRICT HOSPITAL

HERMANN, MISSOURI 65041
(314) 486-2191
MILTON G. WOHLT, ADMINISTRATOR

DIRECTORS: BEN C. RIDDER, PRESIDENT
MARVIN BRUENS, VICE-PRESIDENT
BILLY R. WOLF, TREASURER
ROBERT L. GRANNEMANN
RAYMOND H. KESSLER
EVERETT R. STROTHMANN



June 4, 1981

Mr. D. G. Wiedeman, Acting Chief
Region III
U.S. Nuclear Regulatory Commission
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Dear Sir:

In response to your letter of May 20, 1981 regarding items of noncompliance identified during an inspection of our facility on April 29, 1981, please accept the following actions taken to correct each violation:

- 1) Violation of 10 CFR 20.201 (b). An administrative directive has been issued to the nuclear medicine department director and staff that:
 - a) Whole body film badges and TLD extremity monitors are to be issued to and worn by all employees who prepare and inject patient doses.
 - b) Film and TLD monitors are to be sent for processing at the required monthly intervals.
 - c) Lost or damaged monitors are to be replaced as soon as practicable.
 - d) Film badge reports received showing gaps in exposure coverage for nuclear medicine personnel due to processing errors, lost badges or interruptions in the employment of certain individuals, etc., will be so identified. Estimates of probable exposure based on the average of the three month exposure prior to and the one month following the omitted period will be entered by hand on the film badge report with

8106800474

JUN 8 1981

June 4, 1981

a brief explanation if necessary. The film badge vendor will be requested to update the exposure totals when necessary.

The directive was issued May 4, 1981. My chief technologist has written to our film badge vendor on May 4, 1981 requesting an update of our film badge report and full compliance will be achieved upon receipt of that report.


2) Violation of Condition #14:

An administrative directive has been issued to the director and staff of the nuclear medicine unit to be certain to check the G-M survey meter for constancy prior to each use. The results of this check is to be documented with each recorded survey.

The directive was issued May 4, 1981 and full compliance was achieved with the use of the survey meter on May 4, 1981.

To the best of my knowledge and belief, the statements contained herein are true and correct.

Sincerely,



Milton G. Wohlt
Administrator

MGW/siv