

Detroit Central Hospital

801 VIRGINIA PARK • DETROIT, MICHIGAN 48202 • (313) 874-2800

March 26, 1981

Mr. W. L. Fisher, Chief
Technical Inspection Branch
U.S. Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Re: NRC License No. 21-16892-01
Detroit Central Hospital


Dear Mr. Fisher:

This is in response to our notice of violation on inspection from
January 29, 1981.

1. The Medical Isotope Committee has been formed and will meet at least quarterly. In fact, our first meeting was held two weeks ago.
2. We have ordered from Landauer Company a Finger Tab (TLD) to be used by the apprentice technician in our department.
3. The daily G-M Survey meter will be used to monitor areas used for preparation of patients' dosage on a daily basis.
4. The Texas Nuclear Survey Meter has been sent to Medical Physics Consultants, 3200 West Liberty, Ann Arbor, MI 48104, telephone No. (313) 662-3197 for calibration. This will be performed on a regular basis as specified in our license.

Furthermore, as requested by Ms. Matson, we now have a 250 microcurie source of Cesium - 137, Serial No. 3073MA, which was purchased from Amersham Corporation. I hope this satisfies your requirements.

Sincerely,



Henry A. Shevitz, M.D.
Chief
Nuclear Medicine Department

HAS/jdb

8106300588

MAR 30 1981