



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
799 ROOSEVELT ROAD
GLEN ELLYN, ILLINOIS 60137



JUN 24 1981

Lake of the Ozarks General
Hospital
ATTN: Ms. Bernice Marquart
Director of Finance
P. O. Box 187
Osage Beach, MO 65065

License No. 24-18678-01

Gentlemen:

Thank you for your undated letter received on June 12, 1981, informing us of the steps you have taken to correct the noncompliance which we brought to your attention in our letter dated May 20, 1981.

This letter also references the telephone conversation between Mr. W. P. Reichhold of this office and Mr. James Schaeffer of your staff on June 19, 1981, concerning your corrective actions.

It is our understanding that the following actions will be taken:

1. The Radiation Safety Committee will perform a quarterly review of the following: (a) daily G-M surveys of the elution, preparation, and injection areas, (b) weekly wipe tests of the storage areas, and (c) daily constancy checks of the dose calibrator.
2. New technologists will be instructed in the performance of molybdenum-99 breakthrough tests and a record of this training shall be maintained.
3. The action level for daily G-M surveys of the hot laboratory will be 2 mr/hr and 0.6 mr/hr for all other areas.

IE07
S
//

8107010234 810624
NMS LIC30
24-18678-01 PDR

Lake of the Ozarks General
Hospital

- 2 -

JUN 24 1981

If our understanding of the above is incorrect, please contact
Mr. W. P. Reichhold at 312-932-2500.

We will examine these matters during a future inspection.

Your cooperation with us is appreciated.

Sincerely,

D. G. Wiedeman, Acting Chief
Materials Radiation Protection
Section 1

cc: Fredrick Smeltzer, M.D.
James Schaeffer, R.T., N.M.T.

cc w/ltr rec'd 6/12/81:
DMB/Document Control Desk (RIDS)

RIII *D.V. for*
Reichhold/np
6/23/81

RIII *[Signature]*
Wiedeman
6/23/81