

JUN 23 1981



License No. 24-17477-01

St. Mary's Health Center
ATTN: Mr. Gary R. Stremel
610 West Elm Street
Jefferson City, MO 65101

Gentlemen:

Thank you for your letter dated June 3, 1981, informing us of the steps you have taken to correct the noncompliance which we brought to your attention in our letter dated May 20, 1981.

This letter also refers to the telephone conversation between Mr. W. P. Reichhold of this office and Larry P. Courter, M.D. of your staff on June 19, 1981, concerning your corrective actions.

It is our understanding that the following actions were or will be taken:

1. Daily monitoring of the elution and preparation area and records of the survey results were started on May 1, 1981. A quarterly review of the daily monitoring and survey records will be performed to ensure compliance.
2. Records of the information required by Form NRC-5 will be maintained for new technologists who work with byproduct material.

We will examine these matters during a future inspection.

Your cooperation with us is appreciated.

Sincerely,

D. G. Wiedeman, Acting Chief
Materials Radiation Protection
Section 1

cc: Larry Courter, M.D.
John Lowder, R.T.

cc w/ltr dtd 6/8/81:
DMB/Document Control Desk (RIDS)

IE07
5
1/1

8107010255

OFFICE	RIII <i>for</i>	RIII <i>for</i>			
SURNAME	Reichhold/np	Wiedeman			
DATE	6/23/81	6/23/81			