

APPENDIX A

NOTICE OF VIOLATION

Saint Agnes Medical Center
Philadelphia, Pennsylvania 19145

Docket No.: 30-03196
License No.: 37-13651-01

As a result of the inspection conducted on April 30, 1981, and in accordance with the Interim Enforcement Policy, 45 FR 66754 (October 7, 1980), the following violation was identified:

- A. Condition 16 of your license requires that licensed material be possessed and used in accordance with statements, representations, and procedures contained in your application received May 5, 1980.


Item 15 of this application requires that technologists who administer patient doses containing radioactive material wear disposable gloves.

Contrary to this requirement, as of the day of the inspection, April 30, 1981 disposable gloves were not routinely worn during the administration of patient doses containing radioactive material.

This is a Severity Level V Violation (Supplement VII).

Pursuant to the provisions of 10 CFR 2.201, Saint Agnes Medical Center is hereby required to submit to this office within thirty days of the date of this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Under the authority of Section 182 of the Atomic Energy Act of 1954, as amended, this response shall be submitted under oath or affirmation.

Dated 19 JUN 1981


John D. Kinneman, Chief Materials
Radiological Protection
Section